

Republic of the Philippines  
**BIDS AND AWARDS COMMITTEE**  
 (Goods and Services)  
 City Government of Baguio  
 Baguio City

Standard Form No.: SF-GOODS-60  
 Revised on: September 5, 2008  
 Standard Form Title: Request for Quotation

Reference No.:  
 Shopping Solicitation No.:  
 Date:  
 Ob#::

PR # 104  
 SVP (2) 335  
 March 30, 2017  
 204

Sir/Madam:  
 Please quote your price of item/s listed below as per specifications, stating the shortest time of delivery.  
 Please submit your quotation duly signed by you or your representative to the Bids and Awards Committee not later than 2:00 PM \_\_\_\_\_, 2010 \_\_\_\_\_ c/o the BAC Secretariat, City General Services Office, City Hall, Baguio City. You may fax your quotations through Telefax No. (074)424-5148  
 We may purchase the item/s from your company if the offer is found to be reasonable and advantageous to the City Government of Baguio, but we reserve the right to accept or reject the offer or all quotations found not in order.  
 Thank you very much.

**NOTE:**  
 "PLEASE SEAL YOUR BIDS"

**ATTY LETICIA O. CLEMENTE**  
 BAC Chairperson and City Budget Officer

ITEM NO.	QTY	UNIT	DESCRIPTION	ABC	Bid Price per unit	Bid Price
6	2	packs	<b>DPD Total Chlorine Reagent</b> . 1 pack contains 100 foil pillows of DPD Total Chlorine Reagent . Appearance: white or light pink powder . For measuring bromine and total chlorine of water, wastewater and sewer samples . Range: 0 to 2.00 mg/l . Expiry must be at least 2 years from date of delivery	6,600.00		
7	3	bottles	<b>WATER, DEIONIZED</b> . Capacity: 4 liters/bottle . Dimineralized type: 2 . Specific conductance: ≤ 1 us/cm . Silica ( SiO <sub>2</sub> ): ≤ 30 ug/l . Chloride ( Cl ): ≤ 20 ug/l . Hardness: Negative . Consumption of KNMO . Expiry must be at least 2 years from date of delivery	9,000.00		
8	7	packs	<b>DISSOLVED OXYGEN REAGENT, HIGH RANG</b> . 1 pack contains 25 pieces of high High range Dissolved Oxygen Accuvac ampuls with 2 reusable ampul caps . Range: 0 to 15.00 mg/l O <sub>2</sub> . Appearance: white powder . For measuring Dissolved Oxygen of water and wastewater samples . Expiry must be at least 2 years from date of delivery	22,400.00		
9	1	bottles	<b>pH Buffer Solution</b> . Capacity : 500 - 1000 ml/bottle . Appearance: clear yellow solution . Odorless: infinitely soluble in water . Boiling point : 100°C . Melting point: 0°C . Specific gravit: 1.0 . Vapor density: vapor pressure and evaporation rate is essentially the same as water . For the calibration of multi-parameter meter . Expiry must be at least 1 year from date of delivery	3,000.00		

Purpose : For use by WAMD-CEPMO Baguio Sewage Treatment Plant Laboratory for water and wastewater analysis using a Colorimeter C/O CEPMO

**NOTE:** 1) Delivery Period within \_\_\_\_\_ calendar days