

Republic of the Philippines  
**BIDS AND AWARDS COMMITTEE**  
**(Goods and Services)**  
 City Government of Baguio  
 Baguio City

Standard Form No.: SF -GOODS-60  
 Revised on: September 5, 2008  
 Standard Form Title: Request for Quotation

Reference No:  
 Shopping Solicitation No:  
 Date:  
 CDr#:

PR # 170  
 SVP (2) 340  
 March 30, 2017  
 332

Sir/Madam:  
 Please quote your price of item/s listed below as per specifications, stating the shortest time of delivery.  
 Please submit your quotation duly signed by you or your representative to the Bids and Awards Committee not later than 2:00 PM \_\_\_\_\_, 2010 \_\_\_\_\_ c/o the BAC Secretariat, City General Services Office, City Hall, Baguio City. You may fax your quotations through Telefax No. (074)424-5148  
 We may purchase the item/s from your company if the offer is found to be reasonable and advantageous to the City Government of Baguio, but we reserve the right to accept or reject the offer or all quotations found not in order.

Thank you very much.

**NOTE:**  
**"PLEASE SEAL YOUR BIDS"**

*ATY LETICIA O. CLEMENTE*  
 BAC Chairperson and City Budget Officer

ITEM NO.	QTY	UNIT	DESCRIPTION	Total	Bid Price	Bid Price
				ABC	per unit	
1	1	piece	<p><b>pH Electrode probe</b></p> <p><b>Features:</b></p> <ul style="list-style-type: none"> <li>. Standard digital combination non-refillable, gel-filled pH electrode with epoxy body, double junction Ag/AgCl reference, ceramic pin junction diaphragm, built-in temperature sensor and standard size glass bulb</li> </ul> <p><b>Specifications:</b></p> <ul style="list-style-type: none"> <li>. Measurement range: 0to 14 pH</li> <li>. Temperature range: 0 to 80°C</li> <li>. Junction: Ceramic Pin</li> <li>. Accuracy: ± 0.02pH</li> <li>. Cable length: 1 meter</li> <li>. Diameter: 12mm</li> <li>. Dimensions ( Dxl ) : 12 mm x 200mm</li> <li>. Filling Solution: non-refillable gel</li> <li>. Probe type: Standard</li> <li>. Reference type: Ag/AgCl ( double-junction)</li> <li>. Resolution: 0.001pH</li> <li>. Sensor type: glass</li> <li>. Connector: HQD specific</li> <li>. Product number: Intellical PHC 201</li> </ul> <p>Note:                      Bidder/Supplier must provide an item profile or specifications for evaluation purposes.</p>	25,000.00		

**Purpose :** For WAMD-CEPMO: replacement of pH electrode probe for HACH HQ40d multi-meter  
 C/O CEPMO

**NOTE:**

- 1) Delivery Period within \_\_\_\_\_ calendar days
- 2) Price Validity shall be for a period of \_\_\_\_\_ calendar days ( Minimum of 1 year )
- 3) Warranty a.) For equipment \_\_\_\_\_ ( Minimum of 3 months )  
 b.) For repairs \_\_\_\_\_ ( Minimum of 3 months )

All Bid Prices should be inclusive of VAT and delivery charges