

Republic of the Philippines
BIDS AND AWARDS COMMITTEE
 (Goods and Services)
 City Government of Baguio
 Baguio City

Standard Form No.: SF -GOODS-60
 Revised on: September 5, 2008
 Standard Form Title: **Request for Quotation**

Reference No :
 Shopping Solicitation No :
 Date :
 Obr # :

PR # 101-8731-16-170
SVP (5) 1530
December 7, 2018
101-8731-16-09-332

Sir/Madam:

Please quote your price of item/s listed below as per specifications, stating the shortest time of delivery. Please submit your quotation duly signed by you or your representative to the Bids and Awards Committee not later than 2:00 PM _____ c/o the BAC Secretariat, City General Services Office, City Hall, Baguio City. You may fax your quotations through Telefax No. (074)424-5148.

We may purchase the item/s from your company if the offer is found to be reasonable and advantageous to the City Government of Baguio, but we reserve the right to accept or reject the offer or all quotations found not in order.

Thank you very much.

NOTE:
"PLEASE SEAL YOUR BIDS"

ATTY. LETICIA O. CLEMENTE
 BAC Chairperson and City Budget Officer

ITEM NO.	QTY	UNIT	DESCRIPTION	TOTAL ABC	BID PRICE PER UNIT	TOTAL BID PRICE
1	1	pc	<p>pH Electrode probe</p> <p>Features:</p> <ul style="list-style-type: none"> • Standard digital combination non-refillable, gel-filled pH electrode with epoxy body, double junction Ag/AgCl reference, ceramic pin junction diaphragm, built-in temperature sensor and standard size glass bulb <p>Specifications:</p> <ul style="list-style-type: none"> • Measurement range: 0 to 14 pH • Temperature range: 0 to 80°C • Junction: Ceramic Pin • Accuracy: ± 0.02 pH • Cable length: 1 meter • Diameter: 12 mm • Dimensions: (D x L): 12 mm x 200 mm • Filling Solution: non- refillable gel • Probe type: Standard • Reference Type: Ag/AgCl (double junction) • Resolution: 0.001 pH • Sensor type: glass • Connector: HQD specific • Product number: Intellical PHC 201 <p>Note:</p> <ul style="list-style-type: none"> • Bidder/Supplier must provide an item profile or specifications for evaluation process 	25,000.00		
ac	**Please submit and accomplish the attached Omnibus Sworn Statement together with this form**					

PURPOSE: For WAMD-CEPMO: replacement of pH electrode probe for HACH HQ40d multi-meter. c/o CEPMO

- NOTE:**
- 1) Delivery Period within _____ calendar days
 - 2) Price Validity shall be for a period of _____ calendar days
 - 3) Warranty (for equipments) _____ Minimum of 1 year
 Warranty (for repairs) _____ Minimum of three (3) months
 - 4) Bid Price should be inclusive of VAT and delivery charges
 - 5) Please indicate the brand of your offer. For machines/equipment, please attach brochure.

Name of Establishment	Philgeps Registry No.	Printed Name and Signature of Bidder
Address & Contact Number	page 1 of 1	TIN